Bia Diagnostics Laboratories

Office: 802-540-0148 | Fax: 802-540-0147
480 HERCULES DR. COLCHESTER, VT 05446

Certificate of Analysis

Company: Cloud 9 Canna
Sample ID: Sour Chem
Lot: 0203-10-03
Matrix: Flower
Customer ID: 221129-0
Date Sampled: N/A
Grower License #: SCLT0203
Date Received: 3/26/2024

Report Date: 4/4/2024
Date Analyzed: 4/2/2024
Analyst: 057
Report ID: C240326AJ

Cannabinoid Summary

Cannabinoid Profile	LOQ (mg/g)	Concentration (mg/g)	Weight (%)
CBDVA	0.0005	<LOQ	<LOQ
CBDV	0.0012	<LOQ	<LOQ
CBDA	0.0008	0.78	0.08
CBGA	0.0008	9.75	0.97
CBG	0.0019	1.21	0.12
CBD	0.0019	<LOQ	<LOQ
THCV	0.0021	<LOQ	<LOQ
CBN	0.0013	<LOQ	<LOQ
Δ9-THC	0.0020	20.87	2.09
Δ8-THC	0.0019	<LOQ	<LOQ
THC-A	0.0034	255.18	25.52
CBC	0.0024	<LOQ	<LOQ
Total THC		244.66	24.47
Total CBD		0.69	0.07
Total Cannabinoids		287.79	28.78

24.47% **Total THC**	0.07% **Total CBD**
28.78% **Total Cannabinoids**	2.09% **Δ9-THC**
12.70% **Percent Moisture**	1 : 0 **THC : CBD Ratio**



Cannabinoids Methodology: High Performance Liquid Chromatography (HPLC) using PerkinElmer FLEXAR™ with Photo Diode Array Detector (PDA)

Total CBD and total THC are calculated values, to account for assumed decarboxylation from the acid form (THCA or CBDA) to the neutral form, causing weight loss of the acid group. These values are calculated as follows:
Total THC = (THCA x 0.877) + Δ9-THC
Total CBD = (CBDA x 0.877) + CBD
Ratio of Total CBD: Total THC
Reagent Blanks: < LOQs for all analytes

LOQ = The lowest quantity that this method can reliably detect. Any cannabinoid that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample.

Measurement of Uncertainty (MU): the parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the particular quantity subject to measurement.
Δ9-THC MU = ±0.005%
Total THC MU = ±0.007%

All other cannabinoid MU values are available upon request.

All moisture analysis is determined by loss-on-drying measurement using OHAUS Model MB90 Moisture Content Readers.

Certified by: Luke E. M
Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

Bia Diagnostics Laboratories

Office: 802-540-0148 | Fax: 802-540-0147
480 HERCULES DR. COLCHESTER, VT 05446

Certificate of Analysis

Company: Cloud 9 Canna
Customer ID: 221129-0
Grower License #: SCLT0203

Sample ID: Sour Chem
Lot: 0203-10-03
Matrix: Flower
Date Sampled: N/A
Date Received: 3/26/2024

Report Date: 4/4/2024
Date Analyzed: 3/27/2024
Analyst: 045
Report ID: C240326AJ

Water Activity Summary

Test	Method	Result
Water Activity	ASTM D8196: Determination of Water Activity in Cannabis Flower	0.5420



Test Methodology: Aqualab TDL 2 water activity meter with tunable diode laser

Certified by: Luke E. M
Luke Emerson Mason (Laboratory Director, Bia Diagnostics)